



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

SHINYEI TECHNOLOGY Co., LTD. TSUKUBA CENTER
B47-11 Katoridai Tsukuba, Ibaraki 300-2657

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Calibration of accelerometer on the basis of ISO16063-22
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:
December 28, 2016

Issue Date:
November 29, 2022

Expiration Date:
January 31, 2025

Accreditation No.:
89376

Certificate No.:
L22-796

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*



Certificate of Accreditation: Supplement

SHINYEI TECHNOLOGY Co., LTD. TSUKUBA CENTER

B47-11 Katoridai Tsukuba, Ibaraki 300-2657
Contact Name: Hideki Yamashita Phone: 029-848-3570

Accreditation is granted to the facility to perform the following calibration:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Accelerometer ^F	Charge sensitivity pC/(m/s ²) Accelerometer output ≥ 1 pC 300 m/s ² (pulse width: 11 ms)	2.6 %	"Procedures for Calibration of Accelerometer" (TD-RC-002) On the basis of: ISO16063-22 (Shock) Calibration equipment: Pendulum shock testing machine PST-300 Shock Manager SM-500 Calibration sensor: MEGGITT 2270M8
	1 000 m/s ² (pulse width: 6 ms)	2.2 %	
	10 000 m/s ² (pulse width: 1 ms)	2.3 %	
	20 000 m/s ² (pulse width: 0.5 ms)	2.2 %	
	50 000 m/s ² (pulse width: 0.2 ms)	4.0 %	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represent the smallest measurement uncertainties attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.